

## **Certificate of Analysis**

Product Name: PocketSwab Plus Lot Number: 8032824D

Expiration Date: February 2025

Order Codes: POCK-PLUS

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

| Quality Control Test | Specification        | Result |
|----------------------|----------------------|--------|
| Negative Control     | 0 RLU                | Pass   |
| ATP Positive Control | 10,000 to 40,000 RLU | Pass   |

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg Title: Vice President, Quality Assurance Date: 04/15/2024